Instruction Manual

 $\mathbf{\tilde{\mathbf{C}}}$

TEL: 714-237-9220 FAX: 714-237-9214 1031 Segovia Circle Placentia, CA 92870-7137 USA

MODEL TP-30B MULTIMETER TEMPERATURE ADAPTER

Limited Three Year Warranty

B+K Precision warrants to the original purchaser that its product and the component parts there of, will be free from defects in workmanship and

materials for a period of three years from the data of purchase.

by proof of the purchase date in the form a sales receipt. B+K Precision will, without charge, repair or replace, at it's option, defective product or component parts. Returned product must be accompanied

To obtain warranty coverage in the U.S.A. this product must be registered by completing and mailing the enclosed warranty card to B+K Precision. 1031 Segovia Circle, Placentia, CA 92870 within fifteen (15) days from proof of purchase.

Exclusions : This warranty does not apply in the event of misuse or abuse of the product or as result of unauthorized alternations or repairs. It is vold if the serial number is alternated, defaced or removed.

B+K Precision shall not be liable for any consequential damages, including without limitation damages resulting from loss of use

Some states do not allow limitation of incidental or consequential damages, so the above limitation or exclusion may not apply to you

This warranty gives you specific rights and you may have other rights, which vary from state-to-state





INTRODUCTION

1-1 Unpacking and Inspection

should have the following items: Upon removing your new Thermocouple Module from its packing, you

1. Thermocouple Module

2. K-type Bead Thermocouple

Operator's manual.

1-2 Meter safety

Terms as Marked on Equipment

ATTENTION Refer to manual

+ BATTERY

Symbols In This Manual.

⊳ This symbol indicates where cautionary or other

information is found in the manual.

1-3 Front Panel

with the meter's front panel controls and connectors Refer to Figure 1 and the following numbered steps to familiarize yourself

1. LED : Power on display. (Flashing)

2. Function Switch : Slide switch used to select measuring function

4. Thermocouple Input Connector: This is used to connect the 3. Output Plugs : Voltage output

thermocouple



2-1 General Specifications

National Bureau of Standards and to the IEC 584 Standards for K-type This thermometer conforms to the temperature / voltage tables of the 1. Temperature Scale : Celsius or Fahrenheit user selectable

2. Input : Single K-type thermocouple.

4. Measurement Range :-50°C \sim 1000°C , -58°F \sim 1832°F 3. Output to Meter : 1mV/dc per $^\circ\! C$ or $^\circ\! F$

5. Low Battery Indication :

2. Insert the "output plugs (V and COM)" to the input terminals of 1. Turn to power off

3. Read the display on the multimeter. Replace the battery , if multimeter to "DCV" range.

the reading is less than 7.1V.

6. Temperature Coefficient : 0.1 x (spec.Accuracy) / $^{\circ}F$, < 64 $^{\circ}F$ or >82°F .(0.15 x (spec.Accuracy) / $^\circ \! \mathbb{C}$, <18°C or > 28°C)

7. Power Requirements : Standard 9V battery , NEDA 1604 JIS 006P

or IEC6F22 Size.

number and descrip-	contact name, phone	complete return shipping address,	instrument your co	Include with the ins tion of problem.	t÷ 5
Printed in Taiwan	Facsimile : 714-237-9214	Facsimi			
480-797-9-001	Phone : 714-237-9220	Phone :			
	Placentia, CA 92870	Placentia			
	1031 Segovia Circle	1031 Seg			
	Precision	B+K Pred		¢	
	-	America. For overnight shipments and non-North America shipping fees contact B+K Precision	hipments and non-No	America. For overnight s	≥
tions in North	Return all merchandise to B+K Precision with pre-paid shipping. The flat-rate repair charge includes return shipping to locatic	pre-paid shipping. The fla	o B+K Precision with	Return all merchandise to	
he product.	payment in the form of a money order or credit card. For the most current renair charges contact the factory before shipping the product	t card. For the most curre	money order or credi	payment in the form of a	י כ
ount must include	problem and return any leads, connectors and accessories that you are using with the device. Customers not on open accou	accessories that you an	eads. connectors and	problem and return any le	g :
the performance	Non-Warranty Service : Please return the product in the original packaging to the below address. Clearly state in writing the performance	product in the original pa	ce : Please return the	Jon-Warranty Servic	z 3
ly state in writing the	warrance problem and return any leads, connectors and accessories that you are using with the device.	Dunectors and accessorie	iease letulii tile produ	vallality service : Pi	p 🕈
			- - - - - - -		5
			Decome unstable of In		
	Figure 2	equipment, be aware that	near noise generating of the	If the instrument is used near noise generating equipment, be aware that the voltage output may become unstable or indicate large error.	
	SCREW	neter.	connector of thermom	the thermocouple to the connector of thermometer.	> -
		seconds , after connecting	rarm up for at least 30 s	1. Before measurement , warm up for at least 30 seconds , after connecting	<u>.</u>
		surement	aution before Meas	3-1 Preparation and Caution before Measurement	ų
			OVENS.	MENTS IN MICROWAVE OVENS	ML
A.		ERATURE MEASURE-	O NOT MAKE TEMPE	DAMAGE OR BURNS . DO NOT MAKE TEMPERATURE MEASURE	Ď
		OV DC. TO AVOID	D 24V r.m.s. AC OR 6	MENT SURFACE EXCEED 24V r.m.s. AC OR 60V DC.	M S
4		SHOCK , DO NOT	WARNING : TO AVOID ELECTRICAL SHOCK , DO NOT		
Battery 9V					Ę
	X off	bles for K-type I hermo-	mperature / Voltage tai ermometer	Standards and IEC 584 temperature / Voltage tables for K-type / hermo-	S S
F		ws National Bureau of	erature indication follo	temperature sensor. Temperature indication follows National Bureau of	<u>e</u>
	Ð	pe thermocouple as	ed to use external K-ty	This instrument is designed to use external K-type thermocouple as	7 5
				OPERATION	ç
se. bottom case and screw.	 Calefully disconnect the battery from bottom case. Replace with a new 9V battery and reinstall the bot 	pe insulated.	niature plug. Teflon ta	Wire Length : 1m, with miniature plug. Teflon tape insulated	≦
	3. Remove the bottom case.	$0{}^\circ\!{ m C}$, \pm (4.0 $^\circ\!{ m F}$ or 0.75%)	(2.2°C or 2.0%) from -40°C to 0°C , \pm (4.0°F m -40°E to 32°E	± (2.2℃ c	
			$32^{\circ}\mathrm{F}$ to $399.2^{\circ}\mathrm{F}$.	from 32°F	
meter. from the bottom	nector of the thermocouple from connector of the meter. 2 Position the meter face down Remove the screw from the bottom	04°C , \pm (4.0°F or 0.75%)	\pm (2.2°C or 0.75%) from 0°C to 204°C , \pm	Tolerances : ± (2.2°C or	5
ff. Remove the con-	1. Disconnect the output plugs and turn the meter off. Remove the con-	399 2°F 1	e)°∩ to 204°∩ (-40°∓ to	ວບຣັກ bead thermocouple Temperature Rance : -40°⊂ to 204°⊂ (-40°≖ to 399 2°≖)	H 2
to Figure 7 and use	the following procedure to replace the battery.	rmocouple	eristics:K-Type Ther 2	Thermocouple Characteristics : K-Type Thermocouple	7
rto Figure 3 and use	4-2 Battery Replacement				
nermocouple.	— Avoid chemical reactions that can damage the thermocouple		+(2.9% rda + 4°F)	2	
	—Don't overheat the thermocouple.		±(2.9% rdg + 2℃)	2	
	— Avoid excess bending		±(2.0% rdg + 4°F)	γ	
rmocouple in good	about temperature technology. To maintain a thermocouple in good		±(2.0%rdg+2°C)	351℃ ~ 500℃	
e by a skilled person	ō.		\pm (0.5%rdg + 4° $_{ m F}$)	$-3^{\circ}\mathrm{F}\sim 662^{\circ}\mathrm{F}$	
in a dry cloth and	 I o keep the instrument clean, whether case with detergent, do not use abrasives or solvents. 		±(0.5%rdg + 2℃)	-19° C $\sim 350^{\circ}$ C	
-	4-1 General Maintenance		±(2.0% rg + 4°厈)	$-58^\circ\mathrm{F} \sim -4^\circ\mathrm{F}$	
	LE BEFORE OPENING THE COVER.		<u>+</u> (2.0% rdg + 2°C)	-50°C ∼ -20°C	
VE TEST THERMO-		Input Protection	Accuracy	Range	
	exceeds 60V d.c. or 24V r.m.s.		cations	-J Electrical Specifications	N-3
potential	e surface that the	er.	removed from meter		د
	6. Read the stable reading.	0% R.H. with battery	:-20°C ~ 60°C 0 to 80	. Storage Temperature : -20 $^\circ C \sim 60 ^\circ C$ 0 to 80% R.H. with battery	ω.
	3. Use the sensitily point of the misocouple to measure the surface to be measured.		(35 to 50℃) R.H.	(
		5℃);0% to 70%	1% to 80% R.H. (0 to 3	. Operating Ambient : 0% to 80% R.H. (0 to $35^\circ\mathbb{C}$) ; 0% to	2
				1. Indoor Use.	<u> </u>
of reading will be	COM) of Multimeter. Observe the proper polarity of		onditions	2-2 Environmental Conditions	Ņ
t terminals (V and			operator's manual.	operate	
	tnermometer. 2. Select the °⊂ / °F function desired.	battery (installed).	i ttery) : 114 gms.	 Weight (including battery): 114 gms. Accessories: K - type bead thermocouple _battery 	5 2
	 Connect the plug of the their locadple to the connector of the thornood. 	ע)ו	: 122mm(L) x 46mm(W) x 30mm(D)	9. Dimensions : 122mm(i	<u>ب</u>
ator of the	3-2 remperature measurements		SUU nours.	8. Battery Life : Alkaline 300 hours	°,
					٥