R A Curtis Industries Company

1601 N. CLANCY CT. VISALIA, CA 93291 PH: (559) 651-2222 FAX: (559) 651-0188 http://www.tri-mag.com sales@tri-mag.com

DE300 SERIES Enclosed 300 Watts ITE & Medical, 300W



DESCRIPTION

The DE300 Series is a 300 Watt Enclosed power supply that is 3"x 5"x 2.08" providing 9.6 Watts per cubic inch. Each unit has a built in Active Power Factor Correction and the efficiency of this series is between 89% to 91% depending on model. The DE300 has a built-in forced air cooling and each series has Molex input and output.

FEATURES

- High Efficiency
- Active PFC
- Single Output
- Universal input 90VAC to 264VAC

APPLICATIONS

- IT Applications
- Medical Applications
- Telecommunications
- Test Instrumentation Product
- Data Acquisition
- Other Applications



Operating Temperature

California Efficiency

GENERAL SPECIFICATIONS

Line Voltage	
Input Frequency	47Hz to 63Hz
No load input power	
Inrush Current (cold)	
	or 60A at 230VAC
Operating Temperature	
Storage Temperature	
Cooling	Free Air Convection
	300W 24CFM forced air
Efficiency	
Holdup Time	20ms at 115VAC
Overvoltage Type	
Overload Protection	
	Within 150% rated load
Safety:	
Designed in full compliance	
	EN60950-1
	ANSI/AAMI ES60601-1
	EN60601-1
EMI	
	EN61000-3-3
EMSE	N61000-4-2,-3,-4,-5,-6,-11

MECHANICAL SPECIFICATIONS



All specifications and prices subject to change without notice



1601 N. CLANCY CT. VISALIA, CA 93291 PH: (559) 651-2222 FAX: (559) 651-0188 http://www.tri-mag.com sales@tri-mag.com

OUTPUT SPECIFICATIONS

Model	Watts	Voltage (Vdc)	Load (A)		Voltage	Ripple	Regulation		
			Min.	Rate	Max	Tolerance	& Noise Pk to Pk	Line	Load
DE300-7	300	+12V	0	25	-	+11.9V~+12.1V	120mVpp	±1%	±1%
DE300-8	300	+15V	0	18	-	+14.9V~+15.1V	150mVpp	±1%	±1%
DE300-3	300	+18V	0	16.6	-	+17.9V~+18.1V	180mVpp	±1%	±1%
DE300-9	300	+24V	0	12.5	-	+23.9V~+24.1V	200mVpp	±1%	±1%
DE300-G	300	+28V	0	10.7	-	+27.9V~+28.1V	250mVpp	±1%	±1%
DE300-J	300	+36V	0	8.3	-	+35.9V~+36.2V	250mVpp	±1%	±1%
DE300-14	300	+48V	0	6.3	-	+47.9V~+48.2V	250mVpp	±1%	±1%

Note: Contact factory for Safety Agency Approved status.

- 1. Each output can provide up to peak load temporarily. Continuous operation at greater than rated load is not allowed.
- 2. At factory, in 60% rated load condition, each output is checked to be within voltage accuracy.
- 3. Line regulation is defined by changing $\pm 10\%$ of input voltage from nominal line at rated load.
- 4. Load regulation is defined by changing ±40% of measured output load from 60% rated load.
- The ripple and noise is measured by using 15MHz bandwidth limited oscilloscope. Each output is terminated with a 0.47 μF capacitor at rated load and nominal line.
- 6. Hold up time is measured from the end of the last charging pulse to the time when the main output drops down to 95% output voltage at rated load and nominal line.
- 7. Efficiency is measured at rated load.

MEDICAL ISOLATION GRADE

