

### 65W Single Output Medical Type

# MPS-65 series



#### Features :

- Universal AC input / Full range
- Low leakage current <250µA
- Protections: Short circuit / Overload / Over voltage
- Cooling by free air convection
- Medical safety approved (2 x MOPP between primary to secondary)
- 100% full load burn-in test
- Fixed switching frequency at 45KHz
- 3 years warranty



#### **SPECIFICATION**

MODEL		MPS-65-3.3	MPS-65-5	MPS-65-7.5	MPS-65-12	MPS-65-13.5	MPS-65-15	MPS-65-24	MPS-65-27	MPS-65-48	
OUTPUT	DC VOLTAGE	3.3V	5V	7.5V	12V	13.5V	15V	24V	27V	48V	
	RATED CURRENT	12A	12A	8A	5.2A	4.7A	4.2A	2.7A	2.4A	1.35A	
	CURRENT RANGE	0~15.2A	0~13.8A	0~9.6A	0~6A	0~5.4A	0~4.8A	0~3A	0~2.7A	0~1.5A	
	RATED POWER	39.6W	60W	60W	62.4W	63.45W	63W	64.8W	64.8W	64.8W	
	OUTPUT POWER (max.)	72W(+3.3V:50W;+5V:69W)with 18CFM min. Forced air convection									
	RIPPLE & NOISE (max.) Note.2	80mVp-p	100mVp-p	100mVp-p	100mVp-p	100mVp-p	100mVp-p	100mVp-p	100mVp-p	100mVp-p	
	VOLTAGE ADJ. RANGE	2.97 ~ 3.63V	4.5~5.5V	6.75~8.25V	10.8 ~ 13.2V	12.2 ~ 14.85V	13.5 ~ 16.5V	21.6~26.4V	24.3 ~ 29.7V	43.2~52.8\	
	VOLTAGE TOLERANCE Note.3	±3.0%	±3.0%	±3.0%	±2.0%	±2.0%	±2.0%	±2.0%	±2.0%	±2.0%	
	LINE REGULATION	±1.0%	±1.0%	±1.0%	±1.0%	±1.0%	±1.0%	±1.0%	±1.0%	±1.0%	
	LOAD REGULATION	±3.0%	±3.0%	±3.0%	±2.0%	±2.0%	±2.0%	±2.0%	±2.0%	±2.0%	
	SETUP, RISE TIME	800ms, 30ms	230VAC	800ms, 30ms/*	115VAC at full I	oad					
	HOLD UP TIME (Typ.)	50ms/230VAC 16ms/115VAC at full load									
INPUT	VOLTAGE RANGE	90 ~ 264VAC 127 ~ 370VDC									
	FREQUENCY RANGE	47 ~ 63Hz									
	EFFICIENCY(Typ.)	66%	74%	76%	77%	78%	79%	80%	80%	80%	
	AC CURRENT (Typ.)	1.6A/115VAC	0.9A/230								
	INRUSH CURRENT (Typ.)	COLD START 17A/115VAC 40A/230VAC 40A/230VAC									
	LEAKAGE CURRENT Note.7	Earth leakage current < 250µA/264VAC . Touch current < 60µA/264VAC									
PROTECTION		73 ~ 105W (3.3V:51 ~ 75W)(5V:70 ~ 105W) rated output power									
	VERLOAD Protection type : Hiccup mode, recovers automatically after fault condition is removed										
	OVER VOLTAGE	3.8~4.46V			,	15.5 ~ 18.2V		27.6 ~ 32.4V	31~36.45V	55.2 ~ 64.8	
		Protection tyr									
ENVIRONMENT	WORKING TEMP.	Protection type : Hiccup mode, recovers automatically after fault condition is removed -10 ~ $+60^{\circ}C$ (Refer to "Derating Curve")									
		20 ~ 90% RH non-condensing									
	STORAGE TEMP., HUMIDITY	-20 ~ +85°C, 10 ~ 95% RH									
	TEMP. COEFFICIENT	±0.04%/C (0~50°C)									
	VIBRATION	10 ~ 500Hz, 2G 10min./1cycle, period for 60min. each along X, Y, Z axes									
SAFETY & EMC (Note 4)	SAFETY STANDARDS	ANSI/AAMI ES60601-1, TUV EN60601-1, IEC60601-1 approved									
	ISOLATION LEVEL	Primary-Secondary: 2xMOPP, Primary-Earth:1xMOPP									
	WITHSTAND VOLTAGE	I/P-O/P:4KVAC I/P-FG:2KVAC O/P-FG:0.5KVAC									
	ISOLATION RESISTANCE	I/P-O/P, I/P-FG:100M Ohms / 500VDC / 25°C/ 70% RH									
	EMC EMISSION	Compliance to EN55011 (CISPR11) Class B, EN61000-3-2,-3									
	EMC IMMUNITY	Compliance to EN61000-4-2,3,4,5,6,8,11, EN60601-1-2, medical level, criteria A									
OTHERS	MTBF	359.7Khrs min. MIL-HDBK-217F (25°C)									
	DIMENSION	127*76*42mm (L*W*H)									
	PACKING		s/14.6Kg/1.280	CUFT							
NOTE	<ol> <li>Ripple &amp; noise are measure</li> <li>Tolerance : includes set up</li> <li>The power supply is consid a 360mm*360mm metal pla perform these EMC tests, p</li> <li>Mounting holes M1 and M2</li> <li>Heat Sink HS1,HS2 can nc</li> <li>Touch current was measure</li> </ol>	parameters NOT specially mentioned are measured at 230VAC input, rated load and 25°C of ambient temperature. pple & noise are measured at 20MHz of bandwidth by using a 12" twisted pair-wire terminated with a 0.1 uf & 47 uf parallel capacitor. lerance : includes set up tolerance, line regulation and load regulation. e power supply is considered a component which will be installed into a final equipment. All the EMC tests are been executed by mounting the unit on 160mm*360mm metal plate with 1mm of thickness. The final equipment must be re-confirmed that it still meets EMC directives. For guidance on how to form these EMC tests, please refer to "EMI testing of component power supplies." (as available on http://www.meanwell.com) unting holes M1 and M2 should be grounded for EMI purposes. at Sink HS1,HS2 can not be shorted. uch current was measured from primary input to DC output. oduct Liability Disclaimer : For detailed information, please refer to https://www.meanwell.com/serviceDisclaimer.aspx									



## MPS-65 series

