## CERTIFICATE OF COMPLIANCE

Certificate Number 20151216-E188913
Report Reference E188913-A55-UL
Issue Date 2015-DECEMBER-16

Issued to: TRACO ELECTRONIC AG

SIHLBRUGGSTRASSE 111

CH-6340 BAAR SWITZERLAND

This is to certify that representative samples of

COMPONENT - POWER SUPPLIES, MEDICAL AND

DENTAL

DC-DC converter, THB 6 series.

Have been investigated by UL in accordance with the

Standard(s) indicated on this Certificate.

Standard(s) for Safety: ANSI/AAMI ES60601-1, Medical Electrical Equipment - Part

1: General Requirements for Basic Safety and Essential

Performance:

CSA-C22.2 No. 60601-1, Medical Electrical Equipment -

Part 1: General Requirements for Basic Safety and

Essential Performance.

Additional Information: See the UL Online Certifications Directory at

www.ul.com/database for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's Certification and Follow-Up Service.

Recognized components are incomplete in certain constructional features or restricted in performance capabilities and are intended for use as components of complete equipment submitted for investigation rather than for direct separate installation in the field. The final acceptance of the component is dependent upon its installation and use in complete equipment submitted to UL LLC.

Look for the UL Certification Mark on the product.

Bambles

Bruce Mahrenholz, Director North American Certification Program

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <a href="http://ul.com/aboutul/locations/">http://ul.com/aboutul/locations/</a>

